



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Industrial Scale Company, Inc.
7424 Wallisville Road, Houston, TX 77020

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Dimensional, Mass, Force, and Weighing Devices, and Mechanical Calibration
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:

December 6, 2018

Issue Date:

December 6, 2018

Expiration Date:

February 28, 2021

Accreditation No.:

78692

Certificate No.:

L18-560

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

Industrial Scale Company, Inc.

7424 Wallisville Road, Houston, TX 77020
 Contact Name: Marlon Guillen Phone: 713-672-1846

Accreditation is granted to the facility to perform the following calibrations:

Dimensional

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|--|
| Caliper ^F | Up to 24 in | (120 + 22L) μ in | MicTrack/Gage Blocks |
| Micrometer ^F | Up to 24 in | (29 + 22L) μ in | |
| Indicator ^F | Up to 2 in | (114 μ in + 22L) μ in | |

Mass, Force, and Weighing Devices

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|--|
| Analytical Balances/Scales ^{FO} | 1 mg to 1000 mg | (1.1 x 10 ⁻⁵ + 6 x 10 ⁻⁵ Wt) g | Class F Weights |
| Bench/Counting Scales ^{FO} | 1 lb to 500 lb | (3 x 10 ⁻⁶ + 9 x 10 ⁻⁶ Wt) lb | |
| Floor/ Platform Scales ^O | Up to 10 000 lb | (2.1 + 2.92 x 10 ⁻⁵ Wt) lb | |
| Truck Scales ^O | Up to 120 000 lb | (24 + 2.9 x 10 ⁻⁵ Wt) lb | |
| Force - Compression/Tension ^F | Up to 10 000 lbf | (2.9 + 3 x 10 ⁻⁶ Wt) lb | Dynamometer |
| | 2 000 to 120 000 lbf | 0.22 % of Reading | Load Cell |

Mechanical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|---|--|--|
| Torque Wrench ^F | Up to 50 lbf·in | 0.08 lbf·in + 0.3 % of Reading | Torque Transducer(s) |
| | 50 lbf·in to 400 lbf·in | 0.068 lbf·in + 0.3 % of Reading | |
| | 400 lbf·in to 1 000 lbf·in | 1.7 in·lb + 0.3 % of Reading | |
| | 1 000 lbf·in to 250 lbf·ft | 0.42 lbf·ft + 0.3 % of Reading | |
| | | Up to 600 lbf·ft | 1 lbf·ft + 0.3 % of Reading |
| Pressure Gauges ^F | Up to 30 psi | 0.009 psi + 0.01 % of Reading | Pressure Transducer |
| | 30 psi to 100 psi | 0.062 psi + 0.01 % of Reading | |
| | 100 psi to 200 psi | 0.074 psi + 0.01 % of Reading | |
| | 200 psi to 600 psi | 0.15 psi + 0.01 % of Reading | |
| | 600 psi to 1 000 psi | 0.24 psi + 0.01 % of Reading | |
| | 1 000 psi to 3 000 psi | 0.7 psi + 0.01 % of Reading | |
| | 3 000 psi to 15 000 psi | 3.5 psi + 0.01 % of Reading | |
| | 15 000 psi to 20 000 psi | 160 psi + 0.01 % of Reading | |



Certificate of Accreditation: Supplement

Industrial Scale Company, Inc.

7424 Wallisville Road, Houston, TX 77020

Contact Name: Marlon Guillen Phone: 713-672-1846

Accreditation is granted to the facility to perform the following calibrations:

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this calibration at its fixed location.
4. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations. Example: Outside Micrometer^O would mean that the laboratory performs this calibration onsite at the customer's location.
5. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer^{FO} would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
6. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
7. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.